

✓ To,  
M/s. NIPPEN ELECTRICAL INSTRUMENTS CO.  
12-A, Joy Engg. Compound,  
Marol Maroshi Road,  
Andheri (East), Mumbai-400059

Sub : Approval to use material on works carried out by Government.

Ref : Your letter No. NIL Date:15.12.2006

Sir,

The subject proposal submitted by you was examined by this office and now the undersigned is pleased to accord approval to the following materials for using it on works carried out by Government of Maharashtra W.E.F. the date of issue of this letter.

Name of material	:	<u>EARTH TESTER &amp; MEGGER</u>		
Place of manufacture.	:	12-A, Joy Engg. Compound, Marol Maroshi Road, Andheri (East), Mumbai-400059		
Brand Name	:	<u>NIPPEN</u>		
Technical specifications of the materials	:	1) Earth tester up to 1000 Ohms in single dual range & Triple range with kit as accessories. 2) Megger up to 5000 Volts & Mega Ohms up to 50,000 Mega ohms.		
Manufactured as per ISI No..	:	I.S. 9223	Year-	1991 for (Earth tester)
		I.S. 2992		1987 (Megger)
Whether having ISI No	:	No		
License No.	:	NIL	Validity :	NIL
Whether covered by House hold Electrical Appliances Quality Control orders	:	N.A.		
Manufacturing Certificate No.	:	N.A.		
Validity	:	N.A.		
Authority to issue M.C.	:	N.A.		

The above approval is valid up to JANUARY -2010 provided the performance of the material during this period remains satisfactory, In case of unsatisfactory performance, the approval accorded shall be with drawn forthwith.

Yours faithfully,



(H.B. Raskar)

Chief Engineer (Electrical)